



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

NIOSH Reference: TN-12758
Mfr. Reference: DRA035G

Centers for Disease Control
and Prevention (CDC)
National Institute for Occupational
Safety and Health (NIOSH)
National Personal Protective
Technology Laboratory (NPPTL)
626 Cochran's Mill Road
Pittsburgh, PA 15236
Phone: 412-386-4000
Fax: 412-386-4051

May 30, 2003

Mr. Bodo Heins
Diplom Engineer
Dräger Sicherheitstechnik GmbH
Revalstraße 1
Lubeck 1, D-23560
GERMANY

Dear Mr. Heins:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request dated December 17, 2002. This request was for an extension of approval to add the new p/n 67 38 035 combination cartridge to the X-plore 3300 and X-plore 3500 twin cartridge halfmasks.

NIOSH has reviewed this application and has determined that a new approval number will be granted. Approval is granted for English language only on all documentation. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. Approval number TC-84A-4014 has been assigned. This respirator is approved for protection against organic vapors, chlorine, hydrogen chloride, sulfur dioxide, hydrogen fluoride, formaldehyde, and chlorine dioxide, and P100 particulate efficiency level.

The diskette enclosed with this letter contains the final respirator and cartridge approval labels. The abbreviated label is accepted as submitted. The cautions and limitations which apply to this approval are on the enclosed approval labels. Only those assemblies affected by this request, or where new approval numbers are assigned apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval labels plus the assembly matrix. Parts are to be marked with the numbers indicated on the approval labels in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

Page 2 - Mr. Bodo Heins - TN-12758

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84.

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

Sincerely yours,

A handwritten signature in cursive script, appearing to read "Les Boord".

Les Boord
Acting Branch Chief
Respirator Branch
National Personal Protective Technology Laboratory

Enclosures