



April 18, 2017

Mr. Klaus-Michael Rück
Research & Development of PPE
Dräger Safety AG & Co. KGaA
Revalstraße 1
Lubeck, D-23560
Germany

Dear Mr. Rück:

The National Institute for Occupational Safety and Health (NIOSH) confirms that the following respirators in the Dräger X-plore 1350, 1360 and 1380 series of Filtering Facepieces have been evaluated and are approved for use as Filtering Facepiece Particulate Air Purifying Respirators (APR), in accordance with Title 42, *Code of Federal Regulations*, Part 84 (42 CFR part 84):

TC - Number	Model/Part Number	Description	Protection ¹
TC-84A-5288	3951351	X-plore 1350 N95 Filtering Facepiece, XS	N95
TC-84A-5290	3951368	X-plore 1360 R95 Odor, Filtering Facepiece, S/M	R95
TC-84A-4370	3951352 and 3951353	X-plore 1350 N95, S/M and M/L	N95
TC-84A-4374	3951355 and 3951356	X-plore 1350 N95 V, Filtering Facepiece with Exhalation Valve, S/M and M/L	N95
TC-84A-4375	3951358 and 3951359	X-plore 1350 N95 V Odor, Filtering Facepiece with Exhalation Valve, S/M and M/L	N95
TC-84A-4376	3951362 and 3951363	X-plore 1360 R95 Filtering Facepiece, S/M and M/L	R95

TC-84A-4377	3951365 and 2951366	X-plore 1360 R95 V, Filtering Facepiece with Exhalation Valve, S/M and M/L	R95
TC-84A-7455	3951371 and 3951372	X-plore 1380 N95, Filtering Facepiece, XS and SM	N95

1. Protection:

N95 – Particulate Filter (95% filter efficiency level) effective against particulate aerosols free of oil; time use restrictions may apply.

R95 – Particulate Filter (95% filter efficiency level) effective against all particulate aerosols; time use restrictions may apply.

The approved assemblies consist of the parts as listed on the NIOSH approval labels.

Sincerely,



David Chirdon
Chief, Conformity Verification and Standards
Development Branch
National Personal Protective Technology Laboratory